

Pt10%Rh-Pt (тип „S“) DIN IEC 584

°C	0	1	2	3	4	5	6	7	8	9
0	0.000	0.005	0.011	0.016	0.022	0.027	0.033	0.038	0.044	0.050
10	0.055	0.061	0.067	0.072	0.078	0.084	0.090	0.095	0.101	0.107
20	0.113	0.119	0.125	0.131	0.137	0.142	0.148	0.154	0.161	0.167
30	0.173	0.179	0.185	0.191	0.197	0.203	0.210	0.216	0.222	0.228
40	0.235	0.241	0.247	0.254	0.260	0.266	0.273	0.279	0.286	0.292
50	0.299	0.305	0.312	0.318	0.325	0.331	0.338	0.345	0.351	0.358
60	0.365	0.371	0.378	0.385	0.391	0.398	0.405	0.412	0.419	0.425
70	0.432	0.439	0.446	0.453	0.460	0.467	0.474	0.481	0.488	0.495
80	0.502	0.509	0.516	0.523	0.530	0.537	0.544	0.551	0.558	0.566
90	0.573	0.580	0.587	0.594	0.602	0.609	0.616	0.623	0.631	0.638

°C	0	1	2	3	4	5	6	7	8	9
100	0.645	0.653	0.660	0.667	0.675	0.682	0.690	0.697	0.704	0.712
110	0.719	0.727	0.734	0.742	0.749	0.757	0.764	0.772	0.780	0.787
120	0.795	0.802	0.810	0.818	0.825	0.833	0.841	0.848	0.856	0.864
130	0.872	0.879	0.887	0.895	0.903	0.910	0.918	0.926	0.934	0.942
140	0.950	0.957	0.965	0.973	0.981	0.989	0.997	1.005	1.013	1.021
150	1.029	1.037	1.045	1.053	1.061	1.069	1.077	1.085	1.093	1.101
160	1.109	1.117	1.125	1.133	1.141	1.149	1.158	1.166	1.174	1.182
170	1.190	1.198	1.207	1.215	1.223	1.231	1.240	1.248	1.256	1.264
180	1.273	1.281	1.289	1.297	1.306	1.314	1.322	1.331	1.339	1.347
190	1.356	1.364	1.373	1.381	1.389	1.398	1.406	1.415	1.423	1.432

°C	0	1	2	3	4	5	6	7	8	9
200	1.440	1.448	1.457	1.465	1.474	1.482	1.491	1.499	1.508	1.516
210	1.525	1.534	1.542	1.551	1.559	1.568	1.576	1.585	1.594	1.602
220	1.611	1.620	1.628	1.637	1.645	1.654	1.663	1.671	1.680	1.689
230	1.698	1.706	1.715	1.724	1.732	1.741	1.750	1.759	1.767	1.776
240	1.785	1.794	1.802	1.811	1.820	1.829	1.838	1.846	1.855	1.864
250	1.873	1.882	1.891	1.899	1.908	1.917	1.926	1.935	1.944	1.953
260	1.962	1.971	1.979	1.988	1.997	2.006	2.015	2.024	2.033	2.042
270	2.051	2.060	2.069	2.078	2.087	2.096	2.105	2.114	2.123	2.132
280	2.141	2.150	2.159	2.168	2.177	2.186	2.195	2.204	2.213	2.222
290	2.232	2.241	2.250	2.259	2.268	2.277	2.286	2.295	2.304	2.314

°C	0	1	2	3	4	5	6	7	8	9
300	2.323	2.332	2.341	2.350	2.359	2.368	2.378	2.387	2.396	2.405
310	2.414	2.424	2.433	2.442	2.451	2.460	2.470	2.479	2.488	2.497
320	2.506	2.516	2.525	2.534	2.543	2.553	2.562	2.571	2.581	2.590
330	2.599	2.608	2.618	2.627	2.636	2.646	2.655	2.664	2.674	2.683
340	2.692	2.702	2.711	2.720	2.730	2.739	2.748	2.758	2.767	2.776
350	2.786	2.795	2.805	2.814	2.823	2.833	2.842	2.852	2.861	2.870
360	2.880	2.889	2.899	2.908	2.917	2.927	2.936	2.946	2.955	2.965
370	2.974	2.984	2.993	3.003	3.012	3.022	3.031	3.041	3.050	3.059
380	3.069	3.078	3.088	3.097	3.107	3.117	3.126	3.136	3.145	3.155
390	3.164	3.174	3.183	3.193	3.202	3.212	3.221	3.231	3.241	3.250

Pt10%Rh-Pt (тип „S“) DIN IEC 584

°C	0	1	2	3	4	5	6	7	8	9
400	3.260	3.269	3.279	3.288	3.298	3.308	3.317	3.327	3.336	3.346
410	3.356	3.365	3.375	3.384	3.394	3.404	3.413	3.423	3.433	3.442
420	3.452	3.462	3.471	3.481	3.491	3.500	3.510	3.520	3.529	3.539
430	3.549	3.558	3.568	3.578	3.587	3.597	3.607	3.616	3.626	3.636
440	3.645	3.655	3.665	3.675	3.684	3.694	3.704	3.714	3.723	3.733
450	3.743	3.752	3.762	3.772	3.782	3.791	3.801	3.811	3.821	3.831
460	3.840	3.850	3.860	3.870	3.879	3.889	3.899	3.909	3.919	3.928
470	3.938	3.948	3.958	3.968	3.977	3.987	3.997	4.007	4.017	4.027
480	4.036	4.046	4.056	4.066	4.076	4.086	4.095	4.105	4.115	4.125
490	4.135	4.145	4.155	4.164	4.174	4.184	4.194	4.204	4.214	4.224

°C	0	1	2	3	4	5	6	7	8	9
500	4.234	4.243	4.253	4.263	4.273	4.283	4.293	4.303	4.313	4.323
510	4.333	4.343	4.352	4.362	4.372	4.382	4.392	4.402	4.412	4.422
520	4.432	4.442	4.452	4.462	4.472	4.482	4.492	4.502	4.512	4.522
530	4.532	4.542	4.552	4.562	4.572	4.582	4.592	4.602	4.612	4.622
540	4.632	4.642	4.652	4.662	4.672	4.682	4.692	4.702	4.712	4.722
550	4.732	4.742	4.752	4.762	4.772	4.782	4.792	4.802	4.812	4.822
560	4.832	4.842	4.852	4.862	4.873	4.883	4.893	4.903	4.913	4.923
570	4.933	4.943	4.953	4.963	4.973	4.984	4.994	5.004	5.014	5.024
580	5.034	5.044	5.054	5.065	5.075	5.085	5.095	5.105	5.115	5.125
590	5.136	5.146	5.156	5.166	5.176	5.186	5.197	5.207	5.217	5.227

°C	0	1	2	3	4	5	6	7	8	9
600	5.237	5.247	5.258	5.268	5.278	5.288	5.298	5.309	5.319	5.329
610	5.339	5.350	5.360	5.370	5.380	5.391	5.401	5.411	5.421	5.431
620	5.442	5.452	5.462	5.473	5.483	5.493	5.503	5.514	5.524	5.534
630	5.544	5.555	5.565	5.575	5.586	5.596	5.606	5.617	5.627	5.637
640	5.648	5.658	5.668	5.679	5.689	5.700	5.710	5.720	5.731	5.741
650	5.751	5.762	5.772	5.782	5.793	5.803	5.814	5.824	5.834	5.845
660	5.855	5.866	5.876	5.887	5.897	5.907	5.918	5.928	5.939	5.949
670	5.960	5.970	5.980	5.991	6.001	6.012	6.022	6.033	6.043	6.054
680	6.064	6.075	6.085	6.096	6.106	6.117	6.127	6.138	6.148	6.159
690	6.169	6.180	6.190	6.201	6.211	6.222	6.232	6.243	6.253	6.264

°C	0	1	2	3	4	5	6	7	8	9
700	6.274	6.285	6.295	6.306	6.316	6.327	6.338	6.348	6.359	6.369
710	6.380	6.390	6.401	6.412	6.422	6.433	6.443	6.454	6.465	6.475
720	6.486	6.496	6.507	6.518	6.528	6.539	6.549	6.560	6.571	6.581
730	6.592	6.603	6.613	6.624	6.635	6.645	6.656	6.667	6.677	6.688
740	6.699	6.709	6.720	6.731	6.741	6.752	6.763	6.773	6.784	6.795
750	6.805	6.816	6.827	6.838	6.848	6.859	6.870	6.880	6.891	6.902
760	6.913	6.923	6.934	6.945	6.956	6.966	6.977	6.988	6.999	7.009
770	7.020	7.031	7.042	7.053	7.063	7.074	7.085	7.096	7.107	7.117
780	7.128	7.139	7.150	7.161	7.171	7.182	7.193	7.204	7.215	7.225
790	7.236	7.247	7.258	7.269	7.280	7.291	7.301	7.312	7.323	7.334

Pt10%Rh-Pt (тип „S“) DIN IEC 584

°C	0	1	2	3	4	5	6	7	8	9
800	7.345	7.356	7.367	7.377	7.388	7.399	7.410	7.421	7.432	7.443
810	7.454	7.465	7.476	7.486	7.497	7.508	7.519	7.530	7.541	7.552
820	7.563	7.574	7.585	7.596	7.607	7.618	7.629	7.640	7.651	7.661
830	7.672	7.683	7.694	7.705	7.716	7.727	7.738	7.749	7.760	7.771
840	7.782	7.793	7.804	7.815	7.826	7.837	7.848	7.859	7.870	7.881
850	7.892	7.904	7.915	7.926	7.937	7.948	7.959	7.970	7.981	7.992
860	8.003	8.014	8.025	8.036	8.047	8.058	8.069	8.081	8.092	8.103
870	8.114	8.125	8.136	8.147	8.158	8.169	8.180	8.192	8.203	8.214
880	8.225	8.236	8.247	8.258	8.270	8.281	8.292	8.303	8.314	8.325
890	8.336	8.348	8.359	8.370	8.381	8.392	8.404	8.415	8.426	8.437

°C	0	1	2	3	4	5	6	7	8	9
900	8.448	8.460	8.471	8.482	8.493	8.504	8.516	8.527	8.538	8.549
910	8.560	8.572	8.583	8.594	8.606	8.617	8.628	8.639	8.650	8.662
920	8.673	8.684	8.695	8.707	8.718	8.729	8.741	8.752	8.763	8.774
930	8.786	8.797	8.808	8.820	8.831	8.842	8.854	8.865	8.876	8.888
940	8.899	8.910	8.922	8.933	8.944	8.956	8.967	8.978	8.990	9.001
950	9.012	9.024	9.035	9.047	9.058	9.069	9.081	9.092	9.103	9.115
960	9.126	9.138	9.149	9.160	9.172	9.183	9.195	9.206	9.217	9.229
970	9.240	9.252	9.263	9.275	9.286	9.298	9.309	9.320	9.332	9.343
980	9.355	9.366	9.378	9.389	9.401	9.412	9.424	9.435	9.447	9.458
990	9.470	9.481	9.493	9.504	9.516	9.527	9.539	9.550	9.562	9.573

°C	0	1	2	3	4	5	6	7	8	9
1000	9.585	9.596	9.608	9.619	9.631	9.642	9.654	9.665	9.677	9.689
1010	9.700	9.712	9.723	9.735	9.746	9.758	9.770	9.781	9.793	9.804
1020	9.816	9.828	9.839	9.851	9.862	9.874	9.886	9.897	9.909	9.920
1030	9.932	9.944	9.955	9.967	9.979	9.990	10.002	10.013	10.025	10.037
1040	10.048	10.060	10.072	10.083	10.095	10.107	10.118	10.130	10.142	10.154
1050	10.165	10.177	10.189	10.200	10.212	10.224	10.235	10.247	10.259	10.271
1060	10.282	10.294	10.306	10.318	10.329	10.341	10.353	10.364	10.376	10.388
1070	10.400	10.411	10.423	10.435	10.447	10.459	10.470	10.482	10.494	10.506
1080	10.517	10.529	10.541	10.553	10.565	10.576	10.588	10.600	10.612	10.624
1090	10.635	10.647	10.659	10.671	10.683	10.694	10.706	10.718	10.730	10.742

°C	0	1	2	3	4	5	6	7	8	9
1100	10.754	10.765	10.777	10.789	10.801	10.813	10.825	10.836	10.848	10.860
1110	10.872	10.884	10.896	10.908	10.919	10.931	10.943	10.955	10.967	10.979
1120	10.991	11.003	11.014	11.026	11.038	11.050	11.062	11.074	11.086	11.098
1130	11.110	11.121	11.133	11.145	11.157	11.169	11.181	11.193	11.205	11.217
1140	11.229	11.241	11.252	11.264	11.276	11.288	11.300	11.312	11.324	11.336
1150	11.348	11.360	11.372	11.384	11.396	11.408	11.420	11.432	11.443	11.455
1160	11.467	11.479	11.491	11.503	11.515	11.527	11.539	11.551	11.563	11.575
1170	11.587	11.599	11.611	11.623	11.635	11.647	11.659	11.671	11.683	11.695
1180	11.707	11.719	11.731	11.743	11.755	11.767	11.779	11.791	11.803	11.815
1190	11.827	11.839	11.851	11.863	11.875	11.887	11.899	11.911	11.923	11.935

°C	0	1	2	3	4	5	6	7	8	9
1200	11.947	11.959	11.971	11.983	11.995	12.007	12.019	12.031	12.043	12.055
1210	12.067	12.079	12.091	12.103	12.116	12.128	12.140	12.152	12.164	12.176
1220	12.188	12.200	12.212	12.224	12.236	12.248	12.260	12.272	12.284	12.296
1230	12.308	12.320	12.332	12.345	12.357	12.369	12.381	12.393	12.405	12.417
1240	12.429	12.441	12.453	12.465	12.477	12.489	12.501	12.514	12.526	12.538
1250	12.550	12.562	12.574	12.586	12.598	12.610	12.622	12.634	12.647	12.659
1260	12.671	12.683	12.695	12.707	12.719	12.731	12.743	12.755	12.767	12.780
1270	12.792	12.804	12.816	12.828	12.840	12.852	12.864	12.876	12.888	12.901
1280	12.913	12.925	12.937	12.949	12.961	12.973	12.985	12.997	13.010	13.022
1290	13.034	13.046	13.058	13.070	13.082	13.094	13.107	13.119	13.131	13.143

